

## Supplementary material

**Table S1** Analytical conditions for the detection of analytes in hair samples

LC-MS/MS	Waters ACQUITY UPLC I-Class and Xevo TQ-S			
Column	Imtakt Cadenza CD-C18 HT (150 mm × 2.0 mm, 3 μm)			
Column temp.	40 °C			
Flow rate	0.2 mL/min (0–0.2 min) – 0.3 mL/min (0.2–2 min) – 0.4 mL/min (2–4 min) – 0.5 mL/min (4–5 min)			
Mobile phase	A: 5 mM ammonium acetate + 0.05 % formic acid B: Acetonitrile 5 % B (0–0.2 min) – 20 % B (0.2–2 min) – 50 % B (2–3 min) – 95 % B (3–4.5 min) – 5 % B (4.5–5 min)			
Ionization	Electrospray ionization			
Acquisition mode	SRM			
SRM conditions	Analyte	Monitoring ion ( <i>m/z</i> )	Cone voltage (V)	Collision energy (eV)
	DP	256 > 167	5	15
	DDP	242 > 167	5	15
	LD	235 > 86	30	15
	DLD	207 > 58	4	12
	CP	275 > 230	2	16
	ME	180 > 162	4	12
	CP- <i>d</i> <sub>6</sub>	281 > 230	2	16

Liquid chromatograph-tandem mass spectrometer (LC-MS/MS), selected reaction monitoring (SRM), diphenhydramine (DP), desmethyldiphenhydramine (DDP), lidocaine (LD), desethylidocaine (DLD), chlorpheniramine (CP), methylephedrine (ME)

**Table S2** Analytical validation using spiked 0.4-mm hair segment samples

Analyte	LOD (pg/mg)	LLOQ (pg/mg)	Spiked conc. (pg/mg)	Ac (% bias) <sup>a</sup>	Pc (% CV) <sup>b</sup>	MF (%) <sup>c</sup>	Recovery (%) <sup>c</sup>
DP	5	50	50	-7.4	14.0	88 ± 12	76 ± 11
			1000	1.6	1.9	120 ± 6	100 ± 3
DDP	5	25	50	-11.0	5.7	85 ± 17	88 ± 13
			1000	2.2	3.5	124 ± 10	102 ± 5

LD	5	50	50	-4.8	9.5	103 ± 15	96 ± 4
			1000	-0.3	4.2	114 ± 5	115 ± 7
DLD	5	25	50	-3.5	7.5	117 ± 8	112 ± 8
			1000	1.4	7.1	120 ± 7	123 ± 8

Limit of detection (LOD), lower limit of quantification (LLOQ), accuracy (Ac), precision (Pc), coefficient of variation (CV), matrix factor (MF)

The weights of each 0.4-mm blank hair segment were regarded as 4 µg/segment

<sup>a</sup> Average of five measurements

<sup>b</sup> Calculated based on measurements on five different days

<sup>c</sup> Calculated using 0.4-mm hair segments from six drug-free subjects