

120 intended intraoperative measurements

9 impaired intraoperative measurements due to constant cautery (1), more than one electrode with insufficient skin contact (8, 5 of these consecutively in a single patient)

111 valid intraoperative measurements (93%)

9 measurements with one electrode with insufficient skin contact (8 of these in a single patient)