

Start

Initialization

I:E ratio goal



30-s Wait

Ventilation

Adjustment of PEEP and FiO_2 according to Table 1

Oxygenation goal

Is a pH value entered?

Yes

No

Adjustment of RR and V_T

pH goal

Timer = 10 mins
or $\Delta PEEP \neq 0$
or $\Delta V_T \neq 0$

No

Yes

Adjustment of V_T

Plateau pressure goal

