

**T1 versus T2 weighted Magnetic Resonance Imaging
to Assess Total Kidney Volume
in Patients with Autosomal Dominant Polycystic Kidney Disease**

Journal: Abdominal Radiology

Maatje D.A. van Gastel *, BSc¹; A. Lianne Messchendorp *, MD¹; Peter Kappert, MSc²;
Merel A. Kaatee, BSc^{1,3}; Marissa de Jong, BSc¹; Remco J. Renken, MSc, PhD⁴; Gert J. ter
Horst, MSc, PhD⁴; Shekar V.K. Mahesh, MD² and Ron T. Gansevoort, MD, PhD¹.

On behalf of the DIPAK consortium

Departments of ¹Nephrology, ²Radiology, ³Center for Medical Imaging and ⁴Neuro Imaging
Center, University of Groningen, University Medical Center Groningen, Groningen, the
Netherlands.

* both authors contributed equally to this work

Correspondence: Ron T. Gansevoort

Email: r.t.gansevoort@umcg.nl

Supplementary Table 5. Percentage of approved scans when using T1 or T2 weighted images per MRI scanner.

	T1 [n (%)]	T2 [n (%)]	P value
<i>Magneto Avanto, Siemens</i> ¹	37 (64.9)	51 (89.5)	0.002
<i>Ingenia, Philips</i> ¹	57 (74.0)	70 (90.9)	0.006
<i>GE Medical Systems</i> ¹	49 (75.4)	55 (84.6)	0.19
<i>3 Tesla scanner</i> ²	76 (69.7)	91 (83.5)	0.02
P value	0.6	0.4	

Values represent the number of MR images suitable for volume measurement. P values were calculated using a Chi-square test. 1. 1.5 Tesla scanner; 2. Intera, Philips and Magnetom TRIO, Siemens.